UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/549,623	09/20/2005	Hiroyuki Sugawara	1034509-000001	8092
21839 7590 09/22/2009 BUCHANAN, INGERSOLL & ROONEY PC			EXAMINER	
POST OFFICE	BOX 1404	CECIL, TERRY K		
ALEXANDRIA, VA 22313-1404			ART UNIT	PAPER NUMBER
			1797	
			NOTIFICATION DATE	DELIVERY MODE
			09/22/2009	ELECTRONIC

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

ADIPFDD@bipc.com

	Application No.	Applicant(s)	
	10/549,623	SUGAWARA ET AL.	
Office Action Summary	Examiner	Art Unit	
	Mr. Terry K. Cecil	1797	
The MAILING DATE of this communication ap Period for Reply	pears on the cover sheet with the	correspondence address	
A SHORTENED STATUTORY PERIOD FOR REPL WHICHEVER IS LONGER, FROM THE MAILING D. - Extensions of time may be available under the provisions of 37 CFR 1. after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period. - Failure to reply within the set or extended period for reply will, by statut Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	DATE OF THIS COMMUNICATION 136(a). In no event, however, may a reply be the will apply and will expire SIX (6) MONTHS from the course the application to become ABANDON	DN. imely filed m the mailing date of this communication. IED (35 U.S.C. § 133).	
Status			
Responsive to communication(s) filed on <u>03 ∪</u> This action is FINAL . 2b) Thi Since this application is in condition for allowatelessed in accordance with the practice under	s action is non-final. ance except for formal matters, p		
Disposition of Claims			
4)	awn from consideration.	uirement.	
Application Papers			
9) The specification is objected to by the Examin 10) The drawing(s) filed on is/are: a) accomposed and applicant may not request that any objection to the Replacement drawing sheet(s) including the correct the option of the second se	cepted or b) objected to by the drawing(s) be held in abeyance. So ction is required if the drawing(s) is o	ee 37 CFR 1.85(a). bjected to. See 37 CFR 1.121(d).	
Priority under 35 U.S.C. § 119			
 12) Acknowledgment is made of a claim for foreign a) All b) Some * c) None of: 1. Certified copies of the priority document 2. Certified copies of the priority document 3. Copies of the certified copies of the priority application from the International Bureat * See the attached detailed Office action for a list 	nts have been received. Its have been received in Applica Prity documents have been receive Bau (PCT Rule 17.2(a)).	ition No ved in this National Stage	
Attachment(s) 1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date	4) Interview Summar Paper No(s)/Mail I 5) Notice of Informal 6) Other:	Date	

Application/Control Number: 10/549,623 Page 2

Art Unit: 1797

DETAILED ACTION

Election/Restrictions

1. Restriction is required under 35 U.S.C. 121 and 372. This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

- Group I, claims 1-2, 5-7, drawn to a blood treating set having the technical features of a blood collecting bag and first tube, a blood treating bag and second tube having a filter therein.
- Group II, claims 8, 11-14 and 18-21, drawn to a method of producing a blood treat set, having the technical features of a blood collecting bag and first tube, a blood treating bag and second tube having a filter therein, and the steps of aseptically connecting the tubes using a tube connecting device.
- 2. The inventions listed as Groups I and II do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons: the technical features common between the groups, i.e. two bags with respective tubes connected to each other and a filter intermediate one tube is known in e.g. the prior art of either DE10019985A1 or JP2001276212 of record. Therefore, these features are not *special* technical features and unity is lacking between the groups. Note that the limitations within the apparatus of Group I concerning the how the set is made fail to further structurally define the invention of the completed product.

Art Unit: 1797

3. A telephone call was made to the office of Buchanan, Ingersoll, and Rooney on 9-17-2009 to request an oral election to the above restriction requirement, but did not result in an election being made. Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement may be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention. The election of an invention or species may be made with or without traverse. To preserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).

The examiner has required restriction between product and process claims. Where applicant elects claims directed to the product, and the product claims are subsequently found allowable, withdrawn process claims that depend from or otherwise require all the limitations of the allowable product claim will be considered for rejoinder. <u>All</u> claims directed to a nonelected process invention must require all the limitations of an allowable product claim for that process invention to be rejoined.

Application/Control Number: 10/549,623

Page 4

Art Unit: 1797

In the event of rejoinder, the requirement for restriction between the product claims and the rejoined process claims will be withdrawn, and the rejoined process claims will be fully examined for patentability in accordance with 37 CFR 1.104. Thus, to be allowable, the rejoined claims must meet all criteria for patentability including the requirements of 35 U.S.C. 101, 102, 103 and 112. Until all claims to the elected product are found allowable, an otherwise proper restriction requirement between product claims and process claims may be maintained.

Withdrawn process claims that are not commensurate in scope with an allowable product claim will not be rejoined. See MPEP § 821.04(b). Additionally, in order to retain the right to rejoinder in accordance with the above policy, applicant is advised that the process claims should be amended during prosecution to require the limitations of the product claims. Failure to do so may result in a loss of the right to rejoinder. Further, note that the prohibition against double patenting rejections of 35 U.S.C. 121 does not apply where the restriction requirement is withdrawn by the examiner before the patent issues. See MPEP § 804.01.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Mr. Terry K. Cecil whose telephone number is (571) 272-1138. The examiner can normally be reached on 8:00a-4:30p M-F..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Duane Smith can be reached on (571) 272-1166. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Mr. Terry K. Cecil/ Primary Examiner, Art Unit 1797

tkc